



گزارش آزمون

TEST REPORT



Doc.NO:F-251

Rev.NO:04

Rev.Date:98/05/08

Customer: فولاد زرنده ایرانیان

Cyl.NO: Liq Ar

Cylinder arrival date: 01/08/12

Cylinder size: Ton Tank

Inquiry No: 01L00861-1

Inquiry date: 01/08/12

Test date: 01/08/12

Report No: 01L00861-1

Instrument : GC, O2 An, Dewpoint

Lab:Chemical

#	Component	Concentration	Unit	Accuracy	Test method
1	Methane	≤0.5	ppm	±5%	
2	Oxygen	≤1.5	ppm	±10%	
3	Nitrogen	≤3	ppm	±5%	
4	Moisture	≤3	ppm	±10%	
5	Carbon monoxide	N.D	ppm	±5%	
6	Carbon dioxide	N.D	ppm	±5%	
7	Hydrogen	N.D	ppm	±5%	
8	Argon	balance	

Remark:

A. This test is in accordance with The ISO/IEC 17025 requirements for the competence of testing and calibration laboratories.

NA. This test is not in accordance with The ISO/IEC 17025 requirements for the competence of testing and calibration laboratories./ N.D: Not Detection

This Certificate shall not be copied or reproduced.

This Certificate valid with stamp of farafangas

Uncertainty:

The expanded uncertainty is estimated with attention to coverage factor K=2 normal distribution , providing a level of confidence of approximately 95%

Technical Manager:

Lab Manager:

Date:

Seal:

Farafan Gas Lab:

No,1622,Boustan st. , AbbasAbad Ind. Zone . Km 40 khavaran Rd. , Tehran 3393177669 , iran

Tel:+98 21 22252500
Email: LAB@FARAFANGAS.COM

Fax:+98 21 22226383
www.farafangas.com

